



## Ohio Administrative Code

### Rule 3701:1-58-89 Records of calibration measurements of brachytherapy sources.

Effective: August 15, 2005

---

(A) A licensee shall maintain a record of the calibrations of brachytherapy sources required by rule 3701:1-58-48 of the Administrative Code for three years after the last use of the source.

(B) The record must include:

(1) The date of the calibration;

(2) The manufacturer's name, model number, and serial number for the source and the instruments used to calibrate the source;

(3) The source output or activity;

(4) The source positioning accuracy within the applicators; and

(5) The name of the individual, the source manufacturer, or the calibration laboratory that performed the calibration.

---